



QuickGuide: RealFast™ Variant Detection on StepOne™

Setup for Variant Detection Assays:







- Open the StepOne™ Software (QuickGuide is based on version 2.3) and click **Advanced Setup**. In the Experiment Menu go to **Setup**.
- In **Experiment Properties** define Experiment Name: **Name**
 Instrument: **StepOne™** (48 Wells) or **StepOnePlus™** (96 Wells)
 Type of experiment: **Quantitation – Standard Curve**
 Reagents: **TaqMan® Reagents**
 Ramp speed: **Standard**
- In **Plate Setup** assign **Targets** and **Samples** to selected wells.
 - In **Define Targets** click **Add New Target**, name your targets and assign the correct **Reporter** and **Quencher** (i. e. FAM for targeted variant and VIC for PCR control):

Define Targets			
<input type="button" value="Add New Target"/> <input type="button" value="Add Saved Target"/> <input type="button" value="Save Target"/> <input type="button" value="Delete Target"/>			
Target Name	Reporter	Quencher	Color
HLA-B5701	FAM	NFQ-MGB	
PCR control	VIC	NFQ-MGB	







- Click **Add New Sample** repeatedly to enter all your samples and controls.
- Click **Assign Targets and Samples** and select the total number of wells by click+drag in the **View Plate Layout**.
- Assign targets to selected wells by ticking the **Assign** box in the **Assign target(s) to** field.

» **Note:** Select **None** as passive **reference dye** for Variant Detection RealFast™ Assays (e.g. HLA-B27, HLA-B5701).


- Define your **No Template Control (NTC)**:
 - Select a replicate (2 wells) in the plate layout by click+drag.
 - Select "N" for both targets in the **Task** field.

Assign	Target	Task
<input checked="" type="checkbox"/>	HLA-B5701	  
<input checked="" type="checkbox"/>	PCR control	  

- Define your **Samples, Positive Control** and **Negative Control**:
 - Select a well in the plate layout.
 - Assign sample to selected well by ticking the **Assign** box in the **Assign sample to** field.
 - Select "U" for both targets in the **Task** field.

Assign	Target	Task
<input checked="" type="checkbox"/>	HLA-B5701	  
<input checked="" type="checkbox"/>	PCR control	  

- In **Run Method** select a sample volume of **20 µl** and setup the PCR program:
 - Holding Stage: **3 min at 95°C**
 - Cycling Stage: **40 cycles 15 sec at 95°C and 1 min at 60°C**. Make sure **Data Collection On** is enabled.



- **Save** the experiment.
- Load your PCR tubes / plate and press **START RUN** (green button) to start the run.

Analysis of Variant Detection Assays:

After completing a run or after opening a data file the software displays the Experiment Menu **Analysis:**

- Adjust the **Plot Settings** in the field called **Amplification Plot:**
 - Use plot type: **ΔRn Vs. Cycle**
 - Graph Type: **Linear**
 - Color: **Target**
- Click on the **Analysis Settings** window in the top-right corner:
 - Within the **CT Settings** tab disable the **Automatic Threshold** and adjust the threshold manually above the background signals of the **Negative Control**.
 - Disable **Automatic Baseline**.
Baseline Start Cycle should be set to "3" and **Baseline End Cycle** to "15".
 - Confirm with pressing **Apply Analysis Settings**.

Select a Target

Target	Threshold	Baseline Start	Baseline End
HLA-B5701	0.1	3	15
PCR control	0.1	3	15

CT Settings for the 2 Selected Targets

CT Settings to Use: Use Default Settings

Automatic Threshold

Threshold:

Automatic Baseline

Baseline Start Cycle: End Cycle:

» **Note:** Set threshold value for the FAM channel just above the background fluorescent signal generated by the e.g. HLA-B27 Negative Control. Set the threshold value for the HEX channel at the onset of the exponential phase of the amplification curve«

- Review your samples by selecting individual wells in **View Plate Layout**.
- To show results as table click **View Well Table**.
 - Adjust the table according to your needs by selecting/deselecting the listed features in **Show in Table**.
- To print a report click **Print Report** in the upper menu bar:
 - Select data for the report according to your needs.
- To export results in an Excel or text file click **Export ...** in the upper menu bar:
 - Define **Export Properties** and **Customise Export**.

